

EMS & TRAUMA SYSTEMS

PORLAND STATE OFFICE BUILDING | 800 NE OREGON STREET, SUITE 465, PORTLAND, OREGON 97232-2162



July 9, 2024

To whom it may concern,

The U.S. Food and Drug Administration (FDA) has notified the manufacturer of King Airways that the King LTS-D branded supraglottic airways are not approved for pediatric use. The FDA Warning Letter states that, “Use of this device in a pediatric population raise different questions of safety and effectiveness that may lead to patient adverse events.”

For more information, click on the link to the [FDA Warning Letter](#). For questions, please contact ems.trauma@odhs.oregon.gov.

Sincerely,

A handwritten signature in black ink that reads "David Lehrfeld".

David Lehrfeld, MD
Medical Director
Oregon Health Authority
Public Health Division
Emergency Medical Systems & Trauma Systems Program